



Caution.



Distributor symbol. The symbol is accompanied by the distributor's name and address.



Do not re-use. Indicates a medical device that is intended for one single use only.



Reference number. This symbol must be followed by the reference number for the Aurena product family.



Manufacturer, CE mark, name and full address of manufacturer.  
The CE mark shall be placed next to the manufacturer symbol with the address. CE mark must have a minimum height of 5 mm.



Medical Device



Unique device identifier



Handle with care



Sterilized using irradiation